

Report: Digital Pressure Incliner

s/n: 41631

ECN: 0468369

23 May 2000

Bryan Fabbri
Analytical Services & Materials, Inc.
Hampton, Virginia

SUMMARY

Calibration date: 23 May 2000. Next calibration due: 23 May 2002.

A collection, analysis and calibration of data from a Digital Pressure Incliner, s/n 41631, has been completed. The calibration was performed by the calibration laboratory, Wyle Laboratories. These data were collected by Wyle Laboratories on 23 May 2000.

Model : 361B

Serial Number : 41631

The test data presented in data table format show the deviation test instrument in millibars and true pressure in millibars. The test readings fall within the stated manufacturer tolerance of +/- 0.3 millibars. The calibration prior to adjustment also met the manufacturer tolerance.

Application:

Add corrections to measurements per post calibration table.

